




Water Chemistry (AQUACHECK) Individual Report

AQ580 - (Round 580)

AQ4497 - S.C.A. s.n.c.

Issue Number: 1

Issued: 17/02/2020



Science
for a safer world

LGC Standards Proficiency Testing

1 Chamberhall Business Park | Chamberhall Green | Bury | United Kingdom | BL9 0AP



0001

26 - PFOS and PFOA

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
PFOS	Lab Result	LC-MS/MS	<0.10	µg/L		n/a		N/A	N/A	10	N/A	N/A	N/A	N/A	LC-MS/MS
PFOA	Lab Result	LC-MS/MS	5.45	µg/L	-0.29	5.61	0.08	0.56	N/A	11	5.50	5.63	0.252	0.477	LC-MS/MS