



RIEPILOGO DEI RISULTATI DEL LABORATORIO (LAB348)

Codice Laboratorio	Misurando [U/M]	Tecnica Strumentale	Risultato Laboratorio		Risultati Prova Interlaboratorio		Score (z-score / Z-score)	Segnale	Commenti
			x_i	U_i	X_{pt}	$U_{X_{pt}} (k=2)$			
Lab348	Alaclor [ng/L] ✓	GC-MS	75.0	-	67.2	5.0	26.8	0.6	
Lab348	Aldhn [ng/L] ✓	GC-MS	90.0	-	84.5	8.7	33.8	0.3	
Lab348	Clordano (cis+trans) [ng/L] ✓	GC-MS	70.0	-	94.1	8.2	37.6	-1.3	
Lab348	2,4-DDD [ng/L] ✓	GC-MS	200	-	229	15	92	-0.6	
Lab348	2,4-DDE [ng/L] ✓	GC-MS	60.0	-	68.8	7.2	27.6	-0.6	
Lab348	4,4-DDE [ng/L] ✓	GC-MS	200	-	195	10	78	0.1	
Lab348	2,4-DDT [ng/L] ✓	GC-MS	60	-	108	21	48	-2.0	

OK INSCRITO SU QUALAB